

Certificate of Analysis

PRODUCT : WATER RPE
CODE : 307586
LOT N° : V8F400048F
EXPIRING DATE : 2020/06

METHOD : 6455
EDITION 3

TEST	U.M.	SPECIFICATION	RESULT
Description	-	Clear colourless liquid	Conform
pH at 25°C	-	5 ÷ 7	6
Conductivity	µS/cm	<= 0.1	<0.1
Residue on evaporation	ppm	<= 1	0.4
Residue on calcination	ppm	<= 1	0.1
Oxidizing substances (O)	mg/l	<= 0.4	<0.4
Chloride	ppm	<= 0.1	<0.1
Phosphate	ppm	<= 0.1	<0.1
Nitrate	ppm	<= 0.1	<0.1
Sulphate	ppm	<= 0.1	<0.1
Ag	ppm	<= 0.01	<0.01
As	ppm	<= 0.01	<0.01
Au	ppm	<= 0.01	<0.01
B	ppm	<= 0.01	<0.01
Ba	ppm	<= 0.01	<0.01
Be	ppm	<= 0.01	<0.01
Bi	ppm	<= 0.01	<0.01
Ca	ppm	<= 0.01	<0.01
Cd	ppm	<= 0.01	<0.01
Co	ppm	<= 0.01	<0.01
Cr	ppm	<= 0.01	<0.01
Cu	ppm	<= 0.01	<0.01
Fe	ppm	<= 0.01	<0.01
In	ppm	<= 0.01	<0.01
K	ppm	<= 0.01	<0.01
Li	ppm	<= 0.01	<0.01
Mg	ppm	<= 0.01	<0.01
Mn	ppm	<= 0.01	<0.01
Mo	ppm	<= 0.01	<0.01
Na	ppm	<= 0.1	<0.1
Ni	ppm	<= 0.01	<0.01
Pb	ppm	<= 0.01	<0.01
Si	ppm	<= 0.01	<0.01
Sn	ppm	<= 0.01	<0.01
Sr	ppm	<= 0.01	<0.01
Tl	ppm	<= 0.01	<0.01
V	ppm	<= 0.01	<0.01
Zn	ppm	<= 0.01	<0.01
Zr	ppm	<= 0.01	<0.01

Conform to EN-ISO 3696 grade 3